



Entwicklungs- und Prüflabor
Holztechnologie GmbH
Zellescher Weg 24 · D-01217 Dresden
www.ihd-dresden.de



Durch das Deutsche Akkreditierungssystem Prüfwesen (DAF) nach DIN EN ISO 17025 akkreditiertes Prüflaboratorium
Die Akkreditierung gilt für die nachfolgend aufgeführten Prüfverfahren
Reg. Nr. 111 (DAP-PL-1013/00)

TEST CERTIFICATE

Product: Disposable plastic bags:
230 PA6.6 (with and without print)
160 PA6

Customer: FORMECO S.r.l.
Via Cellini, 33
I-35027 Noventa - Padova

Order: Determination and evaluation of the electrostatic charging of disposable plastic bags for small-sized distillation equipments

Test method: Determination of the electrostatic charging according to EN 13463-1:2002, appendix C;
Evaluation of the charging in accordance with Cenelec Report TR 50404:2003 and the BG-rule 132 (edition 07. 2004)

Test report: No. 275125 from 1. February 2006,
Sample No. 1A, 1B and 2

Test results: Due to the evaluation of the electrostatic chargeability on laboratory conditions the presented transparent plastic bags are classified as electrostatically safe for the use in explosion risk ranges within the zones 0, 1 and 2 for the explosion group IIA.

Dresden, 23. March 2006

Head of laboratory



Engineer in charge